

MICROLAP GAUGES

MANUFACTURER OF SLIP GAUGES AND INSTRUMENTS

5-10-392,INDIRA NAGAR,MUKUNDAWADI,
AURANGABAD -431003 MAHARASHTRA

Mobile no. 9881849358,7385819848 email --- microlap81@gmail.com

GST IN 27ASGPC6680C1ZW

CALIBRATED ON	CERTIFICATE NO.	PAGE	NO.OF PAGES
01/03/2019	MLG/2018/08/004	1	2

CUSTOMER NAME & ADRESS	S. EXPERT ENGINEERS PLOT NO.C-144 MIDC WALUJ Aurangabad - 431136
CALIBRATION LOCATION	LABORATORY
TYPE OF GAUGE	SLIP GAUGES ;T C. M88/1
UNIT UNDER CALIBRATION	T.C.SLIP GAUGES SET SR NO 001
ENVIORNMENTAL CONDITIONS	20°C ± 1°C
REFERENCE MASTER USED	T.C. SLIP GAUGE SET M122/1

NOMINAL LENGTH IN MM	CENTRAL DEVIATION IN μM	NOMINAL LENGTH IN MM	CENTRAL DEVIATION IN μM	NOMINAL LENGTH IN MM	CENTRAL DEVIATION IN μM
1.0005	+0.03	1.06	-0.01	1.21	-0.10
1.001	-0.30	1.07	+0.05	1.22	-0.07
1.002	-0.08	1.08	+0.04	1.23	-0.02
1.003	-0.35	1.09	+0.08	1.24	-0.06
1.004	-0.13	1.10	-0.04	1.25	-0.11
1.005	-0.18	1.11	+0.05	1.26	-0.07
1.006	-0.12	1.12	+0.01	1.27	+0.02
1.007	-0.25	1.13	-0.05	1.28	-0.10
1.008	-0.08	1.14	-0.09	1.29	-0.01
1.009	-0.18	1.15	-0.07	1.30	+0.04
1.01	-0.02	1.16	+0.04	1.31	-0.02
1.02	-0.10	1.17	+0.02	1.32	0.00
1.03	+0.06	1.18	+0.02	1.33	-0.06
1.04	-0.02	1.19	0.00	1.34	-0.06
1.05	-0.05	1.20	+0.05	1.35	-0.06

MICROLAP GAUGES 5-10-392,INDIRA NAGAR, MUKUNDAWADI, AURANGABAD CALIBRATED BY <i>Phan</i>	CHECKED BY <i>Phan</i>
--	---------------------------

NOMINAL LENGTH IN MM	CENTRAL DEVIATION IN μM	NOMINAL LENGTH IN MM	CENTRAL DEVIATION IN μM	NOMINAL LENGTH IN MM	CENTRAL DEVIATION IN μM
1.36	+0.01	0.5	-0.10	7.5	-0.22
1.37	-0.05	1.0	-0.30	8.0	-0.06
1.38	-0.04	1.5	-0.34	8.5	-0.08
1.39	-0.04	2.0	-0.20	9.0	-0.06
1.40	0.00	2.5	-0.22	9.5	-0.24
1.41	-0.12	3.0	-0.45	10.0	-1.35
1.42	-0.12	3.5	-0.24	20.0	+0.01
1.43	-0.02	4.0	-0.22	30.0	-0.10
1.44	-0.10	4.5	-0.16	40.0	-0.80
1.45	-0.06	5.0	-0.31	50.0	-0.83
1.46	-0.08	5.5	-0.15	60.0	+0.01
1.47	-0.06	6.0	-0.13	70.0	-0.10
1.48	-0.04	6.5	-0.34	80.0	-0.18
1.49	-0.04	7.0	-0.10	90.0	-0.22
				100.0	-6.50

MICROLAP GAUGES
5-10-83, P. N. Nagar,
Mukund Nagar, Aurangabad
CALIBRATED BY

Rhem
CHECKED BY