S. EXPERT ENGINEERS

C-144, MIDC Waluj,
Aurangabad - 431136, M.S., India
Mobile: 9225317248 / 7755918690,
Email Id: s.expertengineers11@gmail.com

CALIBRATION REPORT

Description : ANALOG PLUNGER DIAL
Identification No. P.D. 6
Mfg. Sr. No. 1027600

Date of calibration : 13-05-2019
Valid till : 13-05-2020
Certificate No. 6

Sr No.	SLIP SIZE	READING	VARIATION	REMARK
1	1.001	1.001	0.000	OK
2	1.002	1.002	0.000	OK
3	1.003	1.003	0.000	OK
4	1.004	1.004	0.000	OK
5	1.005	1.005	0.000	OK
6	1.006	1.006	0.000	OK
7	1.007	1.007	0.000	OK
8	1.008	1.008	0.000	OK
9	1.009	1.009	0.000	OK
10	1.009	1.009	0.000	OK
11	1.008	1.008	0.000	OK
12	1.007	1.007	0.000	OK
13	1.006	1.006	0.000	OK
14	1.005	1.005	0.000	OK
15	1.004	1.004	0.000	OK
16	1.003	1.003	0.000	OK
17	1.002	1.002	0.000	OK
18	1.001	1.001	0.000	OK

Environmental condition: Temperature 20 ˚ ± 2˚

Method used for calibration:

Equipment/ Reference standard	Identification No.	Calib. date	Valid date
Slip Gauge	Sr.No.- 10526	30-08-2017	30-08-2019

Uncertainty of Measurement: ± 0.5 µm

The reported uncertainty is at coverage factor K=2 which corresponds to a coverage probability of approximately 95% for normal distribution.

Inspected by : V. S. Pingalkar

Approved by : B. R. Kadam